EQUIPMENT CALIBRATION, STANDARDIZATION OR CHECK RECORD

Revised 11/2010, 5/11,9/18

PROCEDURE #6

Date: _____ Checked by: _____

Equipment: COARSE SIEVES E11-17

Previous check date: _____ Next due: _____ Frequency: 12 Months Max

Inspection Equipment and Serial Number:

Caliper readable to 0.0005" (0.01 mm) #: _____

							Action Recommended	
Sieves	**Specification (mm)**	**Serial No. or I.D. No.**	**Reading 1**	**Reading 2**	**Reading 3**	**Average**	**Accept**	**Replace**
3" (75 mm)	72.8 - 77.2							
2 - 1/2" (63 mm)	61.1 - 64.9							
2" (50 mm)	48.5 - 51.5							
1 - 1/2" (37.5 mm)	36.4 - 38.6							
1 - 1/4" (31.5 mm)	30.5 - 32.5							
1" (25.0 mm)	24.2 - 25.8							
3/4" (19.0 mm)	18.4 - 19.6							
5/8" (16.0 mm)	15.5 - 16.5							
1/2" (12.5 mm)	12.11 - 12.89							
3/8" (9.5 mm)	9.2 - 9.8							
5/16" (8.0 mm)	7.75 - 8.25							
1/4" (6.3 mm)	6.1 - 6.5							
#4 (4.75 mm)	4.6 - 4.9							
#5 (4.00 mm)	3.87 - 4.13							