## EQUIPMENT CALIBRATION, STANDARDIZATION OR CHECK RECORD

Revised 07/08, 5/11						
				1	PROCEDURE #70	
Date:			Checked by: _	Checked by:		
Equipment:STRAIGHTEDGE				T 99, T 180, SD 104		
Previous check date: Next due:				Frequency: Region - 12 Mo.		
Field Lab/Area - 12 Mo.  Inspection equipment and serial number:						
Feeler gauge 0.01" thick#:			Glass plate the length of the straightedge			
I.D. Number	Length (10" minimum)	Beveled Edge (one edge min.)	Wide Side Plane to 0.01" Maximum	Cutting Edge Plane to 0.01"  Maximum	Action Taken (none, replaced	
	(Inches)	(Condition)	(Yes/No)	(Yes/No)	or other)	